


यू. ओ. नोट

विषय:—बिड संशोधन अपलोड करवाने के काम में।

विषयान्तर्गत अनुरोध है कि निम्न हस्ताक्षरकर्ता द्वारा जारी की गई बिड क्रमांक F 6()/JDAg/QC/2017/1080 Dated 08.02.17 में किये गये संशोधन उपरांत संशोधित प्रपत्र (संलग्न प्रपत्र) ई-प्रोक, एसपीपीपी एवं विभागीय वैबसाईट पर अपलोड करवाने का श्रम करावें।

संलग्न—उक्तानुसार


(के.सी. मीना)
संयुक्त निदेशक कृषि (गु.नि.)

एसीपी
कृषि निदेशालय, जयपुर

यू0ओ0 नोट क्रमांक एफ 6 () नि.कृ./गु.नि./2016-17 / 7559

दिनांक 17-2-17

Government of Rajasthan

Office of Joint Director of Agriculture (QC) Durgapura, Jaipur

No.- F 6()/JDAg/QC/2017/ 1126

Date:- 17.02.17

Corrigendum

With reference to NIB No. F 6()/JDAg/QC/2017/1080 Dated 08.02.17 the modifications/additions specified below will superceed all references made to this regard in the following sections of request for proposal (RFT)-supply and installation of laboratory equipments from manufacturers or their authorized distributors/dealers for the use in pesticide and bio-pesticide testing laboratory under Deptt. of Agriculture, Govt. of Rajasthan.

A. Points Related to Tender Document:

Item No.-1 GLC

| S. No. | Page No./Point | Section No./Original Clause | Modified/Additional Clause | Compliance |
|--------|----------------|--|--|------------|
| 1 | 60 | Point no. 2- Injector Point no. 3- Capillary Injector | Point no. 3 Deleted. Total 3 Injector required (2 Packed+1 Capillary) | - |
| 2 | 62 | Point no. 6-FID (Dual Type) | Dual Type Deleted, 3 Detectors Required | |
| 3 | 62 | Point no. 6A-FID Type:-Quartz Jet 0.32 mm id | FID Type:- Quartz/SS Jet 0.32 mm id | |

Rest terms and conditions will remain same in the bid document.

³
17-2-17
(K. C. Meena)

Joint Director Agriculture (QC)

Durgapura, Jaipur

E-mail- jdagr_qc@rediffmail.com